



# PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidanan, Nabunturan, Davao de Oro

## TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 24-C1684

Bid Number: B-24-0285  
Date: 10/14/2024  
Page: 1 of 2

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
1	2.0	unit	Defibrillator, Brand New, Branded  TECHNICAL SPECIFICATIONS  Power Supply Input: 220 - 240 VAC, 60 Hz, three-pronged plug with built in rechargeable battery pack Biphasic Technology Digital display Output Energy : 0 - 360 joules adjustable or higher Discharge Modes: Synchronous/Asynchronous (Manual) & AED operation (with voice prompt)  Features: <ul style="list-style-type: none"> <li>• Charge control buttons on control panel and on APEX paddle</li> <li>• Charge energy selection on control panel and on APEX paddle</li> <li>• Charging indicators</li> <li>• Isolated ECG monitoring via defibrillator paddles</li> <li>• ECG monitoring via patient cable with sync button</li> <li>• Port for external paddles &amp; AED</li> <li>• Defib can be used both adult &amp; children</li> <li>• Can perform discharge TEST with the unit</li> <li>• With built in printer</li> <li>• With pulse oximeter (SpO2)</li> </ul> Alarm Indicators (Audible and/or Visual): <ul style="list-style-type: none"> <li>• Battery Condition Charging</li> <li>• Low Battery</li> <li>• High and Low Heart Rate</li> <li>• Lead contact failure</li> </ul> Standard Accessories: (per unit) <ul style="list-style-type: none"> <li>• Set of External Paddles ( Adult/Pedia)- Manual Defib - 1 set</li> <li>• AED Pads - 3 sets for Adult &amp; 3 sets for Pedia</li> <li>• ECG patient cables - 1 set</li> <li>• Reusable SpO2 probe - 1 set</li> <li>• Ground cables - 1 set</li> <li>• Power cord - 1 set</li> <li>• Dust cover</li> <li>• Accessory bag</li> <li>• 10 rolls/ z-fold recording paper; size 63 mm width maximum</li> <li>• Medical Grade Mobile cart with brake</li> </ul>	358,499.00	716,998.00		
SUB-TOTAL					716,998.00		
Purpose: FOR THE USE OF DDOPH-MARAGUSAN (DEFIBRILLATOR) 4TH QUARTER Delivery: PGSO Period of Delivery: 10 DAYS							



# PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidanan, Nabunturan, Davao de Oro

## TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 24-C1684

Bid Number: B-24-0285  
Date: 10/14/2024  
Page: 2 of 2

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
			<ul style="list-style-type: none"> <li>• Appropriate AVR(25% higher than the equipment rated value), 220 V, 60 Hz, servo motor type; with over-current protection</li> </ul> <p>TERMS AND CONDITION:</p> <ol style="list-style-type: none"> <li>1. Certification that they will submit Original copy of Operator/Technical Manual in English language.</li> <li>2. Original brochure or manual (from the manufacturer) that contains all the technical specifications required by the bidding documents in English language : attached in the bidding documents.</li> <li>3. The bidder must provide Service Unit which the end user can use in case the equipment will be pulled out for repair / maintenance within the warranty period.</li> <li>4. Certification that the bidder will provide the training for at least two (2) end users and two (2) for the maintenance staff on site.</li> <li>5. Certification that the bidder shall be responsible for the notification, transportation, delivery, installations and commissioning at no cost to the government.</li> <li>6. Two (2) years warranty for parts and services. Warranty period starts/commence after the installation and commissioning, end user's training on operation and acceptance by the end-user/hospital authorized personnel.</li> <li>7. Certification of free preventive maintenance visits(includes check-up, calibration verification and functionality test reports): semi -annually within the warranty period and provide reports.</li> <li>8. The bidder must provide information and technician mobile number to contact anytime.</li> </ol>				
<b>GRAND TOTAL</b>					716,998.00		
Purpose: FOR THE USE OF DDOPH-MARAGUSAN (DEFIBRILLATOR) 4TH QUARTER Delivery: PGSO Period of Delivery: 10 DAYS							